



ATLANTIC SUPPLY

QUALITY DRILLING SUPPLIES AND
MATERIALS TESTING EQUIPMENT

4187 S.W. 34th Street
Orlando, FL 32811
Phone: 800-569-8950 Fax: 407-246-6518

Verification Certificate For Field Thermometer Used To Measure The Temperature of Asphalt Mixtures

Company Name	ABC Material Testing Associates, Inc.	
Company Address	123 Any Street Anytown, FL 12345	
Calibration Date	3/19/12	
Next Calibration Due	3/19/13	
Thermometer Type	Dial	
Model Number	82550DGC	
Temperature Range	50 °F - 550 °F	
Divisions	5 °F	
Serial Number	AS74082	
Calibrated By:	David Thompson	

Thermometer Verification

Calibration Chamber Temperature °F	Field Thermometer Actual Reading °F	Correction Factor °F
75.00	75	0
175.00	175	0
250.00	250	0
350.00	350	0
450.00	450	0

Calibration Equipment:

Circulating Water Bath: Polyscience Dual Bath Thermometry Calibration System, Temperature Range -40 °C - 200 °C, (-40 °F - 392 °F)

Programmable/Digital Controller: P/N 9712A11C, S/N 108300846

Thermometry Calibration System: P/N 040512, S/N C07200315

Fluid: Propylene glycol inhibited

Reference Thermometer: RTD Platinum Thermometer, Control Company, Model #4132, Serial # 80216640, NIST Traceable,

Certificate #8448-64553, Procedure # 3-SCP-004, Temp. Range -58 °F to 752 °F (-50 °C to 400 °C), Resolution is 0.01° below 200° and 0.1° above 200°, Accuracy is +/- (0.1% + 0.2 °C) below 200 °C and +/- (0.15% + 0.5 °C) above 200 °C, Cal. Date: 3-12-12

Dry Well: Hart Scientific, Model # 9100S-B, "B" Block, Serial # A84137, Temp. Range 95 °F - 707 °F (35 °C - 375 °C), Temperature Display and Setability Resolution is 0.1 Degrees, Divisions 0.1 °F or °C, NIST Traceable, Cal. Date: 3-13-12

This certificate shall not be considered or regarded as any type of guarantee or warranty of any form, in favor of the user, the user's customers, or the public at large, that this instrument will continue to retain the same percentage of accuracy as reported on this date, when the calibration, verification, and/or adjustment was performed by Atlantic Supply. Atlantic Supply has absolutely no control over the future operation, performance, damage, maintenance repairs, cleaning and overall condition of said instrument and hereby expressly disclaims any and all liability for damage or loss sustained by any and all parties arising from deterioration, obsolescence, malfunction or sub-standard performance of this instrument. Liability for maintaining this instrument in the same operating condition as it was found to be in on the date of this certificate is the sole responsibility of said instrument's owner.

This form may not be reproduced in any form whatsoever, except in whole, without the expressed written permission of Atlantic Supply